

Trip Report - Nassawadox Grid Site Visit June 12, 2013

Wednesday, 12 June 2013 PB

NASA0017 (C2A) Hospital Water Treatment Plant

1000 – 1022 local

Gauge A was significantly off level, Gauge B was slightly off level.

Cleaned and leveled gauges and solar panel as needed. No problems found.

NASA0007 (D1) Major Jones Farm

1029 – 1035 local

Both gauges were level.

No problems found.

NASA0015 (E1) Campbell Airfield

1039 – 1045 local

Gauge A was level. Gauge B was slightly off level. Leveled gauges as needed. Did not need cleaning.

NASA0008 (E2) Gladden Farm

1053 – 1110 local

Gauge A was level but the tipping assembly had dropped out of the pivot points. Adjusted gauge A pivot points as needed. Gauge B was slightly out of level. Leveled gauge B as needed. No other problems found.

NASA0013 (C3) Dairy Farm

1110 – 1120

Both gauges were level. Gauge B tipping assembly had come out of its pivot points. Adjusted gauge B as needed. No other problems found.

NASA0014 (D4) TNC HQ

1127 – 1131

Both Gauges were level. No problems found.

Note: This trip was done because of anomalies that occurred with the above sites on two rain events subsequent to the June 4 – 6 scheduled maintenance which is listed separately.